

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 1 – Product Identification

This 'Product Safety Data Information' Sheet covers Pall Medical disposable filter cartridges, each employing a PTFE membrane with a polyester support layer filter medium within an acrylic body.

Example Product name(s): Intervene

Example Part Number(s): See appendix 1

The filters detailed above are intended for patient protection in medical applications.

- Gas Insufflation in Minimal Invasive Surgery – MIS
- Filtration of Medical Gases for example Oxygen, Carbon dioxide (CO₂)
- Vacuum Protection

For further information on Pall products, please visit Pall at <https://www.pall.com/en/about-pall.html>

SECTION 2 - Hazards Identification

Product definition: Article.

These products are not classified as hazardous according to REACH Regulation 1907/2006, or European CLP/GHS Regulation 1272/2008.

GHS Signal word: No signal word.

Hazard statements: No known significant effects or critical hazards.

Special packaging requirements: None.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 3 - Materials of Construction

3.1 The filters detailed in Section 1 are comprised of the following materials:

Material Name	CAS Number
Modified acrylic copolymer housing body	Polymer supplier proprietary information
Polytetrafluoroethylene (PFTE) membrane	9002-84-0
Polyester non-woven support layer	25038-59-9

These products are not known to contain BADGE, NOGE, or BFDGE.

Trace additives will be present in the plastic components.

There are no additional ingredients present which, within the current knowledge of the supplier, are classified and contribute to the classification of the article.

There are no current SVHC substances known to be present in the finished articles above 0.1%.

There are no current ROHS2 Directive (2011/65/EU) and amendment (2015/863) substances of concern (including Lead, Cadmium, Mercury, Hexavalent Chromium, Polybrominated biphenyl (PBB), Polybrominated diphenyl ether (PBDE), Bis(2-ethylhexyl) phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl phthalate (DBP) and Diisobutyl phthalate (DIBP)) known to be present in the materials employed in excess of the limits laid down, based on information from our suppliers and knowledge of substances used within Pall the manufacturing facility.

Pall Medical filters do not employ natural rubber latex, or latex derivatives in their construction.

These products (see appendix 1) do not contain animal materials (i.e. animal parts, tissues, or body fluids). However, to assist our customers in performing a TSE/BSE risk assessment, we are pleased to provide the following information:

Pall filters for healthcare and laboratory applications may utilise components which are fabricated from plastic resins containing animal-derived additives at trace levels such as tallow-derived substances. Please be advised that tallow-derived additives are not considered specified BSE risk materials according to the current revision of the U.S. Code of Federal Regulations, Title 21 of part 189.5. Furthermore, the CPMP's *Note for guidance on minimising the risk of transmitting animal spongiform encephalopathies via human and veterinary medicinal products* (EMA410/01 rev 3) gives specific consideration to tallow derivatives and state they are unlikely to be infectious due to the rigorous processing steps used during their manufacture (an example of which is transesterification, or hydrolysis, at not less than 200 °C under pressure for not less than 20 minutes).

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 4 - First Aid Measures

4.1 First aid measures

Always address any contaminants present on the filter as the result of use.

Eye Contact:	Eye injury could result from physical impact. Get medical attention immediately.
Inhalation:	Inhalation is not considered a likely route of exposure for the filter product as supplied by Pall.
Skin Contact:	Wash with soap and water. If irritation persists, get medical attention.
Ingestion:	This material is not intended for ingestion and is not expected to present an ingestion hazard in the form and quantities present in a work place setting. However if ingestion occurs, seek medical attention.
Protection of first-aiders:	No action shall be taken involving any personal risk or without suitable training.

4.2 Key symptoms and effects

No known significant effects or critical hazards related to the materials of construction of the filter as supplied.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 5 - Fire Fighting Measures

5.1 Extinguishing media

Select an extinguish medium suitable for surrounding / working environment.

For filter alone use dry chemical, CO₂, water spray (fog) or foam.

5.2 Specific Hazards

Warning: Combustion products of PTFE (fluoropolymers) can be released and be hazardous to humans and the environment.

Hazardous thermal decomposition products: CO, CO₂, acrid smoke, SO_x, Benzenesulfonic acid, 2(or 4) methyl- phenol,

Warning: thermal decomposition of PTFE can also produce fume particles, toxic gases, hydrofluoric acid and carbonyl fluoride.

Polymer fume fever – chills, nausea, shortness of breath, chest tightness, muscle or joint ache – seek immediate medical attention.

Irritation to eyes. - suitable PPE and breathing apparatus precautions should be taken related to this risk in the event of fire.

5.3 Advice to Fire Fighters

Special precaution required. Fire-fighters should wear appropriate protective equipment, including self-contained breathing apparatus (SCBA) with a full face-piece operated in positive pressure mode.

Protective gloves must be worn when handling debris after a fire, due to PTFE thermal decomposition risks.

SECTION 6 - Accidental Release Measures

Warning: Do NOT incinerate without additional consideration of risk emissions and residues resulting from combustion of PTFE – please see Section 13.

6.1 Personal precautions, protective equipment and emergency procedures

No special measures are required in respect of the filters in the unused condition as supplied.

For used filters always address any contaminants present on the filter as the result of use.

6.2 Environmental precautions

For unused filter modules, place in designated waste container appropriate to the materials of construction listed in Section 3 and dispose of in accordance with local regulations via a licenced waste disposal contractor.

Pall Corporation provides the information contained herein in good faith and states that it represents the best information currently available. However no warranties or representations are expressed or implied and Pall Corporation assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Corporation be responsible for losses or damages resulting from the use or reliance on this information.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

For used filter modules, using clear-up, containment and appropriate PPE measures related to the product being filtered and the materials of construction detailed in Section 3.

6.3 Spillage containment and cleaning up

Use suitable equipment to collect the filter material and place in a designated, labelled waste container.

Care should be taken to consider the nature of any contamination on the filter as the result of use and suitable PPE employed for handling medical waste.

Dispose of waste via a licensed waste disposal contractor.

SECTION 7 – Handling and Storage

7.1 Handling

Put on appropriate personal protective equipment for the working environment (See Section 8). Consult details of product being filtered for specific advice. Avoid activities that can damage the filter.

Follow good hygiene practices. Eating, drinking and smoking are generally prohibited in areas where this product is handled, stored or processed – exceptions are made on the guidance of local medical advice. Staff must follow standard work-place hygiene before eating, drinking or smoking after using this product. Wear gloves to prevent contamination of the filter cartridge and maintain cleanliness of the unused filter.

7.2 Storage

In the received condition, special protective equipment is not needed during handling and normal use of these filters. However, gloves are recommended to prevent contamination of the filter and maintain cleanliness. Handling of used filters must take into account the nature of potential contaminants.

The article is supplied dry, without the presence of any preserving fluid.
Store in clean, dry conditions suitable for a medical device.

Handle with care to avoid damage.

Do not expose to direct sunlight during storage, or other radiation or direct weather conditions.
Store in original shipping bag or boxing.

Ensure careful handling to avoid physical damage. Ensure shipping bag and seals are intact prior to use - do not use if damaged.

Please also consult Pall for further instructions for use information on the product prior to use.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 8 - Exposure Controls/Personal Protection**8.1 Control parameters**

Occupational Exposure limits: None required.

Recommended monitoring procedures: None required

8.2 Exposure controls

There are no special ventilation requirements for the article as supplied in the new and unused condition.

Hygiene Measures: No special measures required. Good hygiene practice in line with local working environmental requirements and medical guidelines.

Hand protection: Disposable gloves are recommended to ensure filter remains clean during installation.

Environmental Exposure Controls: Not normally required for the filter itself as supplied.

After the filter has been used additional exposure controls care should be taken in line with the nature of any contaminant on the filter as a result of its use.

SECTION 9 - Physical and Chemical Properties

Appearance: Disposable filter

Physical state: Solid

Colour: Clear / White outer housing with white filter material

Solubility: Insoluble in water

Acrylic components readily soluble in esters, ketones and chlorinated hydrocarbons

Auto-ignition temperature: Acrylic components: 440 °C (830 °F), decomposition begins at 250 °C (482 °F)

PTFE components: 520° - 560 °C, thermal decomposition >300 °C

Polyester components > 175 °C

Sensitive to shock: Mechanical / thermal shock can result in damage to the filter

Pall Corporation provides the information contained herein in good faith and states that it represents the best information currently available. However no warranties or representations are expressed or implied and Pall Corporation assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Corporation be responsible for losses or damages resulting from the use or reliance on this information.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 10 – Stability and Reactivity

Reactivity:	The filter is stable under the recommended conditions of use and storage.
Chemical Stability:	The filter is stable under recommended conditions of use and storage.
Hazardous Polymerisation:	Polymerisation will not occur under recommended conditions of use and storage.
Other hazardous reactions:	Consult details of product being filtered for specific advice. Under normal conditions of storage and use, no hazardous reactions will occur.
Conditions to Avoid:	Avoid conditions that soften, swell or adversely affect the filter or its materials of construction. Do not allow fluids to freeze on the filter.
Incompatible Materials:	Strong Acids, Strong Alkalis, Strong Oxidising Agents.
Decomposition Products:	Under recommended conditions of use or storage, no hazardous decomposition products will be produced.

SECTION 11 - Toxicological Information

The information in this section contains generic advice and guidance in respect of the unused filter as supplied. Consult SDS details of the product being filtered for specific advice and recommendations.

11.1 Acute Toxicity

Based on typical information for the material type named, this information has not been determined specifically for Pall Medical filters:

Material Name	Result	Species	Dose	Exposure
PTFE	LD50	Rabbit	7.88 g/kg	dermal

Irritation/Corrosion/Sensitisation: No known concern to unused filter as supplied

Mutagenicity / Carcinogenicity / Reproductive Toxicity / Teratogenicity: **No known concern for the materials of construction of the filter as supplied (new and unused)**

Aspiration Hazard: Not applicable for un-used filter.

Potential acute health effects: **No known significant effects or critical hazards for the unused filter as supplied.**

Pall Corporation provides the information contained herein in good faith and states that it represents the best information currently available. However no warranties or representations are expressed or implied and Pall Corporation assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Corporation be responsible for losses or damages resulting from the use or reliance on this information.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

11.2 Chronic health effects

No known significant effects or critical hazards for the unused filter as supplied.

Carcinogenicity: **No specific test data available, no evidence for hazardous properties**

SECTION 12 - Ecological Information

Pall Medical filters are not expected to degrade in contact with soil or water under ambient conditions.

SECTION 13 - Disposal Information

The information in this section contains generic advice and guidance.

Product

Methods of disposal:

Unused as supplied filters: Disposal/handling of the un-used filters should be in-line with national legislation and local regulatory requirements for the materials present. Unused filter cartridges may be used as land-fill.

Warning: Do NOT incinerate unused filters with general waste, as combustion products of PTFE (fluoropolymers) can be released and be hazardous to humans and the environment.

Hazardous Waste: To the best of our knowledge, this product if unused is not regarded as hazardous waste as defined by the EU Directive 91/689/EEC and amendments.

Used filter cartridges should be disposed of as clinical waste due to the nature of the contaminants on the filters as a result of use. Therefore used filters may be classified as hazardous – clinical waste.

Packaging

Bagging: Plastic (polyethylene)

Box: Cardboard

The generation of waste should be avoided or minimised wherever possible. Waste packaging should be recycled where suitable arrangements and facilities exist. Incineration or land-fill should only be considered where re-cycling is not feasible.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

SECTION 14 - Transport Information

The clean and un-used filter, supplied in its original packaging, is not classified as dangerous goods under ADR, RID, IMDG or IATA regulations.

Date of issue: 17 December 2019

Version: 2

Notice to Reader

To the best of our knowledge, the information contained herein is accurate. However, neither the above Pall Corporation, nor any of its subsidiaries, assumes any liability whatsoever for the accuracy or completeness of the information contained herein.

Final determination of suitability of any materials is the sole responsibility of the user. All materials may present unknown hazards and should be used with caution. Although certain hazards are described herein, we cannot guarantee that these are the only hazards that exist.

Pall, and Intervene™ are a trademark of the Pall Corporation

® indicates a trademark registered in the USA

© Pall Corporation, 2019.

PRODUCT SAFETY DATA INFORMATION

Date: 17 December 2019

Data Sheet Number: PSDS2019PMED06 Revision: 2

APPENDIX 1

PALL PART NUMBER	PRODUCT FAMILY	PRODUCT DESCRIPTION
8001022	Intervene™	1 µm PTFE with 6.4 mm (1/4 in.) HB inlet and outlet connectors
8002035	Intervene™	1 µm PTFE with 22 mm (0.9 in.) Female Inlet and 9.5 – 12.7 mm (3/8 -1/2 in.) stepped HB Outlet connectors.
6044197	Round Filter Acrylic-Gas	25MM 0.2UM PTFE/POLYESTER NON-WOVEN MALE/MALE
6124197	Round Filter Acrylic-Gas	25MM .45 PTFE BULK GENERIC
6194197	Round Filter Acrylic-Gas	25MM DLL W/0.45 UM PTFE
6284197	Round Filter Acrylic-Gas	25MM 0.45 PTFE, ACRYLIC MLS/MLS, BULK NS
6484197J	Round Filter Acrylic-Gas	25MM D.L.L.FILTER 0.2 PTFE/ POLYESTER NON-WOVEN, 50/BX, DLL
6544192	Round Filter Acrylic-Gas	25MM 0.2UM PTFE/ POLYESTER NON-WOVEN NON-STERILE

Pall Corporation provides the information contained herein in good faith and states that it represents the best information currently available. However no warranties or representations are expressed or implied and Pall Corporation assumes no liabilities resulting from its use. Users should make their own investigations to determine the suitability of the information for their applications. In no way shall Pall Corporation be responsible for losses or damages resulting from the use or reliance on this information.